

touchTymp MI 36

Tympanometry & Audiometry in one device

Simply intuitive middle ear diagnostics with AC & BC audiometry

Improve your workflow every day

Our touchTymp MI 36 combines diagnostic middle ear analyzer and audiometry tests in one exceptional device. The full 10.4" touchscreen with its user-friendly interface guarantees a unique intuitive operation and allows to quickly change parameters. You can easily switch between the tests without reorganizing from one to another device and benefit from its small footprint when facing limited desk space. touchTymp MI 26 really improves your workflow every day.

Comprehensive test protocols

Our touchTymp MI 36 provides comprehensive standard protocols for immediate operation of screening and diagnostic test processes: tympanometry with 226 Hz, tympanometry with automatic or fixed acoustic reflexes, reflex decay, Eustachian tube function testing (ETF) plus air conduction and bone conduction audiometry. The optional high-frequency probe tone of 1 kHz is ideal for providing reliable results when testing newborns.

Stay focused on your patients

The unique touchTymp light bar on our shoulder box probe provides a real-time progression of the immittance test. The shoulder box can easily be clipped to your patient's top for controlled handling of the diagnostic probe. The light indicator shows you which ear is tested and the test result.

Printing made easy

Print test results instantly by using the touchTymp built-in printer. Place the shoulder box into its holder to automatically start printing. You can transfer the results to a PC to allow long term storage, full page printout and PDF creation for EMR integration.

Features at a glance

- Combo device with small footprint
- Customizable to suit individual needs
- High resolution and quick screen transitions
- Full touch-based interface for intuitive handling
- Probe tone 226 Hz, 678 Hz, 800 Hz
- Optional high-frequency probe tone 1 kHz
- Acoustic reflex stimuli 500 Hz, 1 kHz, 2 kHz, 4 kHz, BB, HP, LP (ipsilateral and contralateral)
- Reflex Decay test (ipsilateral and contralateral)
- ETF test for intact and perforated eardrums
- Air conduction audiometry
- Bone conduction audiometry
- Built-in printer or data transfer to PC



Technical Data touchTymp MI 36

TYMPANOMETRY

Probe Frequency	226 Hz \pm 1 %, 85 dB SPL \pm 1.5 dB 678 Hz \pm 1 %, 72 dB SPL \pm 1.5 dB 800 Hz \pm 1 %, 70.5 dB SPL \pm 1.5 dB
Optional High Frequency	1 kHz \pm 1 %, 69 dB SPL \pm 1.5 dB
Pressure Range	- 600 to + 400 daPa
Accuracy of Pressure	\pm 5 % or \pm 10 daPa
Volume Range	0.0 to 6.0 ml (compensated)
Compliance Range	0.1 to 8.0 ml at 226 Hz 0.1 to 15.0 mmho at 678, 800 and 1000 Hz
Accuracy of Volume	\pm 5 % or 0.1 ml
Test Time Measurement	3 - 5 seconds

ACOUSTIC REFLEXES

Test Frequencies	0.5, 1, 2, 4 kHz \pm 1 %
Test Noise	BB, LP, HP
Test Methods	Ipsilateral, contralateral
Level Ipsilateral	70 to 105 dBHL
Level Contralateral	70 to 120 dBHL
Level Setting	Automatic, fixed
Ipsilateral Reflex Test	With AGC

REFLEX DECAY

Standard	Probe frequency 226 Hz
Test Frequencies	0.5, 1, 2, 4 kHz \pm 1 %
Test Noise	BB, LP, HP
Level Ipsilateral	70 to 105 dBHL
Level Contralateral	70 to 120 dBHL

EUSTACHIAN TUBE FUNCTION

Test Methods	Intact and perforated
Pressure Range	- 600 to + 400 daPa


Standard Components



touchTymp
with printer



Shoulder-Box



DD45 and B71

Optional Accessories



DD45 C
(Tympometry)

AUDIOMETRY

Test Signals	Sinus and warble tone (pulsed and continuous)
Test Frequencies	125, 250, 500, 750, 1000, 1500, 2000, 3000, 4000, 6000, 8000 Hz (125 Hz excluded from bone conduction test frequencies)
Level Steps	5 or 1 dB
Hearing Level Range AC	-10 to 120 dBHL
Hearing Level Range BC	-10 to 80 dBHL
Masking Noise	Narrow band and white noise

DEVICE GENERAL

Display	10.4" Graphic LED-Display with resistive touchscreen
PC Interface	USB
Probe	Lightweight hand-held probe with built-in control light and switch
Printer	Fast 4 inch thermal printer
Power Supply	Mains 100 to 240 V \sim \pm 10 %, 50 - 60 Hz \pm 10 %
Dimensions / Weight	W 30 x D 34.5 x H 14.8 cm / 3.2 kg
Languages	English, German, Spanish, French, Polish, Turkish, Russian, Chinese

STANDARD

- ANSI/AAMI ES/ IEC/ EN 60601-1, class I, Type B; IEC 60645-5, Type 2/
ANSI S3.39, Type 2; IEC 60645-1, Type 3/ ANSI S3.6, Type 3
- Class IIa according to EU medical device directive 93/42/EEC



STANDARD COMPONENTS

Shoulder-Box probe, Shoulder-Box adapter kit, Shoulder-Box attachment kit, contra phone (CIR55), DD45 headset, B71 Bone Conductor, patient response switch, mic-monitor headset, built-in calibration cavity, built-in printer and roll of thermal paper, mains cable, Sanibel ear tip kit, probe floss kit, MAICO USB flash drive kit, MAICO software, USB-cable, cleaning cloth, touchpen, operational manual, quick guides

OPTIONAL ACCESSORIES

Audiometric transducer (DD45 C), talk back microphone

SANIBEL

We highly recommend to use Sanibel disposables in order to guarantee optimal test results.



Specifications are subject to change without notice.

MAICO Diagnostics GmbH

Sickingenstr. 70-71 · 10553 Berlin · Germany
Tel.: +49 30 / 70 71 46-50 · Fax: +49 30 / 70 71 46-99
sales@maico.biz · www.maico.biz

представител за България
Илан Медицинска Апаратура ООД
София, бул. „Джеймс Баучер“ № 76А
Hill Tower Business Center et. 1
тел. 070017373
E-mail: office@ilan.bg Web: www.ilan.bg